

Supplement

Title:	374, 375, 376 Users	Supplement Issue:	5
Part Number:	3608883	Issue Date:	10/12
Print Date:	July 2010	Page Count:	2
Revision/Date:			

This supplement contains information necessary to ensure the accuracy of the above Users Manual.

Change #1

On page 3, under **Warning** add the following:

- **When batteries are changed, ensure that the calibration seal in the battery compartment is not damaged. If damaged, the Meter may not be safe to use. Return the Meter to Fluke for replacement of the seal.**

Change #2, 56750

On page 27, under Operating Altitude:

Change:	Operating Altitude	3000 meters
To:	Operating Altitude	2000 meters



Change #3

On page 4, add:

- **Measure a known voltage to verify product operation.**

Change #4, 63774

On page 7, add the following to the **Symbols** table:

	Conforms to relevant South Korean EMC Standards.
	Conforms to CAN/CSA-C22.2 No. 61010-1, second edition, including Amendment 1.

On page 27, add the following to EMC:

Applies to use in Korea only.....Class A Equipment (Industrial Broadcasting & Communication Equipment) ^[1]

[1] This product meets requirements for industrial (Class A) electromagnetic wave equipment and the seller or user should take notice of it. This equipment is intended for use in business environments and is not to be used in homes.

On page 28, replace **Agency Approvals** with:

Agency Approvals.....